
Midmark M11 - v1.0.5
Total Cycles: 4037

Name: RITTER2
Sterilizer ID: V2313228
10/17/2024 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 150.6 F 1.6
4:00 166.0 F 4.4
6:00 221.9 F 8.4
8:00 207.4 F 13.1
10:00 219.6 F 6.5
12:00 229.7 F 10.9
14:00 232.2 F 7.5
16:00 244.9 F 13.2
18:00 242.7 F 11.9
20:00 254.6 F 18.0
22:00 265.3 F 24.3
24:00 270.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.0 F 29.6
2:00 273.0 F 29.6
3:00 271.4 F 29.0
4:00 271.8 F 29.9

Min 271.2 F 29.0
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 234.5 F 0.0
10:00 183.2 F 0.0
15:00 179.0 F 0.0
20:00 175.6 F 0.0
25:00 173.8 F 0.0
30:00 174.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

10/17/2024 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____