
Midmark M11 - v1.0.5
Total Cycles: 3438

Name: RITTER2
Sterilizer ID: V2313228
03/07/2024 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.1 F 0.0
2:00 135.5 F 0.2
4:00 136.3 F 2.0
6:00 158.9 F 5.2
8:00 183.3 F 9.4
10:00 200.7 F 8.9
12:00 220.3 F 7.7
14:00 231.9 F 12.4
16:00 233.0 F 9.3
18:00 239.0 F 10.0
20:00 245.1 F 13.5
22:00 256.7 F 19.7
24:00 267.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 271.5 F 29.6
2:00 272.7 F 29.9
3:00 271.5 F 29.3
4:00 271.2 F 29.5

Min 271.0 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 228.5 F 0.0
10:00 227.7 F 0.0
15:00 229.7 F 0.0
20:00 228.5 F 0.0
25:00 231.0 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:24
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

03/07/2024 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____