
Midmark M11 - v1.0.5
Total Cycles: 3890

Name: RITTER2
Sterilizer ID: V2313228
08/23/2024 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.3 F 0.0
2:00 105.4 F 0.0
4:00 127.6 F 0.7
6:00 157.9 F 3.7
8:00 185.7 F 7.3
10:00 201.4 F 11.8
12:00 217.3 F 5.0
14:00 229.0 F 9.5
16:00 234.3 F 10.5
18:00 239.1 F 11.4
20:00 237.8 F 9.7
22:00 250.7 F 15.9
24:00 261.5 F 22.2
26:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 273.0 F 29.8
2:00 271.5 F 29.4
3:00 272.7 F 29.4
4:00 272.1 F 29.8

Min 271.4 F 29.1
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 222.8 F 0.0
10:00 219.5 F 0.0
15:00 183.9 F 0.0
20:00 177.8 F 0.0
25:00 176.3 F 0.0
30:00 175.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 27:10
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:05:43
CYCLE COMPLETE

08/23/2024 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____