

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3601

Name: RITTER2  
Sterilizer ID: V2313228  
05/09/2024 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.1 F 0.0  
2:00 81.7 F 0.0  
4:00 121.9 F 0.0  
6:00 104.2 F 1.0  
8:00 139.9 F 3.7  
10:00 166.4 F 7.2  
12:00 191.9 F 10.9  
14:00 222.2 F 4.9  
16:00 223.2 F 7.2  
18:00 232.2 F 11.0  
20:00 231.1 F 7.3  
22:00 241.0 F 11.8  
24:00 237.3 F 9.6  
26:00 250.2 F 15.7  
28:00 261.1 F 21.9  
30:00 271.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 29.2  
1:00 272.8 F 29.7  
2:00 272.0 F 29.8  
3:00 272.0 F 29.1  
4:00 273.0 F 29.6

Min 271.6 F 29.0  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.1 F 0.0  
5:00 225.1 F 0.0  
10:00 227.8 F 0.0  
15:00 206.4 F 0.0  
20:00 226.7 F 0.0  
25:00 238.9 F 0.0  
30:00 206.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 31:13  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:20  
CYCLE COMPLETE

05/09/2024 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_