
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006787
14:55 27/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:22
STERILIZING 3:20
136.5 C 222kPa 3:20
136.0 C 220kPa 3:50
136.2 C 222kPa 4:20
136.0 C 217kPa 4:50
136.4 C 222kPa 5:20
136.4 C 220kPa 5:50
136.4 C 223kPa 6:20
136.4 C 223kPa 6:50
136.4 C 223kPa 6:50
VENTING 6:50
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
74D9209E84EB57A8
