
Midmark M11 - v1.0.0
Total Cycles: 4756

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2024 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.3 F 0.0
2:00 117.0 F 1.3
4:00 134.7 F 3.9
6:00 158.7 F 6.9
8:00 192.5 F 10.5
10:00 206.8 F 6.6
12:00 215.3 F 6.3
14:00 226.9 F 10.2
16:00 235.3 F 9.0
18:00 236.8 F 10.4
20:00 239.3 F 8.6
22:00 246.2 F 13.5
24:00 256.9 F 19.1
26:00 266.7 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.0 F 29.0
2:00 272.4 F 29.0
3:00 271.5 F 28.4
4:00 271.9 F 29.1

Min 271.2 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 200.2 F 0.0
10:00 196.7 F 0.0
15:00 195.1 F 0.0
20:00 196.1 F 0.0
25:00 201.4 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:38
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:06:27
CYCLE COMPLETE

07/22/2024 04:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____