
Midmark M11 - v1.0.0
Total Cycles: 5248

Name: RITTER 1
Sterilizer ID: V2178198
01/16/2025 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.5 F 0.0
4:00 84.8 F 0.0
6:00 115.7 F 0.0
8:00 142.9 F 0.2
10:00 170.7 F 0.5
12:00 195.8 F 0.7
14:00 212.8 F 3.6
16:00 227.2 F 6.7
18:00 235.7 F 10.1
20:00 237.1 F 9.8
22:00 240.1 F 10.6
24:00 235.4 F 7.9
26:00 243.8 F 11.5
28:00 246.0 F 13.0
30:00 258.6 F 19.8
32:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.6 F 28.8
2:00 272.4 F 28.4
3:00 271.6 F 27.9
4:00 272.4 F 28.7

Min 271.6 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.1 F 0.0
5:00 210.1 F 0.0
10:00 226.2 F 0.0
15:00 201.6 F 0.0
20:00 205.3 F 0.0
25:00 233.7 F 0.0
30:00 225.4 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:11:19
CYCLE COMPLETE

01/16/2025 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____