

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4293

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2025 11:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.8 F 0.0  
2:00 130.2 F 1.1  
4:00 176.4 F 3.5  
6:00 198.0 F 7.0  
8:00 188.8 F 10.9  
10:00 206.5 F 3.9  
12:00 219.8 F 6.3  
14:00 230.7 F 10.4  
16:00 233.3 F 8.4  
18:00 241.0 F 11.6  
20:00 238.8 F 10.0  
22:00 251.0 F 16.0  
24:00 261.7 F 22.1  
26:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.7 F 29.5  
2:00 272.0 F 28.8  
3:00 271.3 F 28.9  
4:00 272.4 F 29.4

Min 271.1 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 222.2 F 0.0  
10:00 191.0 F 0.0  
15:00 185.9 F 0.0  
20:00 181.8 F 0.0  
25:00 179.5 F 0.0  
30:00 177.7 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:41  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:05:28  
CYCLE COMPLETE

01/24/2025 01:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_