

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2896

Name: RITTER 1  
Sterilizer ID: V2178198  
02/14/2023 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
2:00 196.5 F 3.1  
4:00 195.2 F 7.5  
6:00 212.2 F 12.7  
8:00 222.7 F 7.4  
10:00 235.7 F 12.7  
12:00 237.2 F 9.9  
14:00 236.4 F 8.4  
16:00 249.0 F 14.9  
18:00 261.2 F 21.8  
20:00 272.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.1  
2:00 271.5 F 28.4  
3:00 272.7 F 29.2  
4:00 271.9 F 28.6

Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 220.8 F 0.0  
10:00 210.7 F 0.0  
15:00 229.0 F 0.0  
20:00 226.7 F 0.0  
25:00 228.9 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:20  
CYCLE COMPLETE

02/14/2023 11:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_