
Midmark M11 - v1.0.5
Total Cycles: 4291

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 07:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.9 F 0.0
2:00 141.6 F 2.3
4:00 164.2 F 6.1
6:00 191.1 F 10.8
8:00 217.7 F 4.6
10:00 229.8 F 9.7
12:00 234.8 F 9.1
14:00 241.8 F 12.2
16:00 240.4 F 10.8
18:00 253.0 F 17.2
20:00 264.1 F 23.8
22:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.5
1:00 272.2 F 29.6
2:00 272.2 F 29.5
3:00 272.3 F 29.5
4:00 272.7 F 29.5

Min 272.1 F 29.4
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 219.2 F 0.0
10:00 183.7 F 0.0
15:00 183.0 F 0.0
20:00 179.8 F 0.0
25:00 179.0 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

01/23/2025 08:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____