
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002196
12:06 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:57
STERILIZING 2:55
136.0 C 221kPa 2:55
135.8 C 222kPa 3:25
136.4 C 223kPa 3:55
136.3 C 225kPa 4:25
136.4 C 226kPa 4:55
136.4 C 226kPa 5:25
136.1 C 224kPa 5:55
136.1 C 224kPa 6:25
136.1 C 224kPa 6:25
VENTING 6:25
AIR DRYING 7:19
CYCLE COMPLETE 13:50

Digital Signature #
17409F9323FFABBB
