

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2867

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2023 09:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
2:00 183.1 F 3.2  
4:00 185.0 F 7.2  
6:00 205.0 F 12.0  
8:00 217.1 F 5.8  
10:00 231.4 F 11.1  
12:00 232.2 F 7.8  
14:00 242.9 F 11.8  
16:00 244.7 F 12.6  
18:00 257.6 F 19.5  
20:00 269.1 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.9  
2:00 271.6 F 28.6  
3:00 272.5 F 29.0  
4:00 271.5 F 28.4  
  
Min 271.5 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 229.8 F 0.0  
10:00 237.3 F 0.0  
15:00 230.7 F 0.0  
20:00 228.1 F 0.0  
25:00 227.1 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 21:03  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:20  
CYCLE COMPLETE

02/07/2023 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_