

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2695

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2023 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 157.5 F 0.2  
4:00 165.9 F 3.2  
6:00 187.2 F 6.8  
8:00 204.0 F 10.7  
10:00 215.3 F 5.3  
12:00 226.7 F 8.6  
14:00 237.0 F 13.2  
16:00 238.2 F 10.6  
18:00 236.1 F 8.8  
20:00 249.4 F 15.3  
22:00 260.2 F 21.4  
24:00 270.0 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.4 F 29.6  
2:00 271.6 F 28.8  
3:00 271.9 F 29.5  
4:00 272.3 F 29.3

Min 271.2 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 212.5 F 0.0  
10:00 219.9 F 0.0  
15:00 216.5 F 0.0  
20:00 216.2 F 0.0  
25:00 210.7 F 0.0  
30:00 210.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:03:35  
CYCLE COMPLETE

07/25/2023 10:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_