
Midmark M11 - v1.0.5
Total Cycles: 3448

Name: RITTER2
Sterilizer ID: V2313228
03/08/2024 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.8 F 0.0
2:00 142.6 F 1.3
4:00 154.2 F 4.2
6:00 180.2 F 8.0
8:00 198.9 F 12.3
10:00 213.9 F 5.7
12:00 229.0 F 10.2
14:00 231.7 F 9.0
16:00 241.5 F 12.7
18:00 242.0 F 11.7
20:00 254.5 F 18.1
22:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.8 F 29.8
2:00 272.3 F 29.9
3:00 271.7 F 29.5
4:00 271.9 F 29.3

Min 270.8 F 28.6
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 229.5 F 0.0
10:00 206.6 F 0.0
15:00 204.1 F 0.0
20:00 200.9 F 0.0
25:00 198.2 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:29
CYCLE COMPLETE

03/08/2024 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____