

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002262  
13:25 20/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:34  
STERILIZING 3:30  
136.0 C 220kPa 3:30  
136.2 C 224kPa 4:01  
136.1 C 222kPa 4:31  
136.1 C 218kPa 5:01  
136.2 C 222kPa 5:31  
136.3 C 222kPa 6:01  
136.2 C 218kPa 6:31  
136.4 C 223kPa 7:01  
136.4 C 223kPa 7:01  
VENTING 7:01  
AIR DRYING 7:55  
CYCLE COMPLETE 15:36

-----  
Digital Signature #  
3A78EA17E93C8009  
-----