
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 005072
14:45 06/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:08
STERILIZING 4:11
136.5 C 225kPa 4:11
136.4 C 225kPa 4:41
136.3 C 224kPa 5:11
136.2 C 223kPa 5:41
136.1 C 219kPa 6:11
136.4 C 225kPa 6:41
136.2 C 219kPa 7:11
136.1 C 223kPa 7:41
136.1 C 223kPa 7:41
VENTING 7:41
AIR DRYING 8:37
CYCLE COMPLETE 9:10

Digital Signature #
15B4A9BCAF2E4A69
