
Midmark M11 - v1.0.5
Total Cycles: 4285

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 71.9 F 0.0
4:00 79.2 F 0.1
6:00 113.4 F 1.7
8:00 143.7 F 4.7
10:00 171.1 F 8.0
12:00 189.8 F 11.6
14:00 202.7 F 3.2
16:00 217.0 F 6.4
18:00 227.5 F 10.1
20:00 231.1 F 10.0
22:00 237.9 F 10.8
24:00 235.1 F 8.5
26:00 248.3 F 14.8
28:00 259.0 F 20.8
30:00 268.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.9 F 29.6
2:00 271.8 F 28.9
3:00 272.9 F 29.7
4:00 271.9 F 28.8

Min 271.7 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 191.0 F 0.0
10:00 185.9 F 0.0
15:00 189.0 F 0.0
20:00 189.4 F 0.0
25:00 188.3 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:10
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

01/23/2025 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____