

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3596

Name: RITTER2  
Sterilizer ID: V2313228  
05/07/2024 02:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.2 F 0.0  
2:00 83.7 F 0.0  
4:00 94.3 F 0.1  
6:00 128.9 F 1.2  
8:00 151.4 F 3.6  
10:00 169.8 F 6.4  
12:00 186.2 F 9.6  
14:00 200.3 F 12.7  
16:00 213.8 F 5.6  
18:00 221.4 F 9.2  
20:00 230.1 F 13.2  
22:00 233.8 F 9.7  
24:00 238.3 F 10.4  
26:00 243.6 F 12.9  
28:00 254.1 F 18.3  
30:00 263.6 F 24.0  
32:00 272.5 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.6 F 30.2  
2:00 272.1 F 29.7  
3:00 271.2 F 29.4  
4:00 272.2 F 30.2  
  
Min 271.2 F 29.3  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 219.4 F 0.0  
10:00 199.7 F 0.0  
15:00 195.1 F 0.0  
20:00 200.9 F 0.0  
25:00 208.4 F 0.0  
30:00 213.3 F 0.0  
  
Drying Complete

FILLING: 0:55  
HEATING: 32:30  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:11:08  
CYCLE COMPLETE

05/07/2024 03:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_