
Midmark M11 - v1.0.0
Total Cycles: 5066

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2024 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 158.5 F 0.3
4:00 166.3 F 1.5
6:00 191.0 F 5.3
8:00 207.7 F 9.3
10:00 220.5 F 13.4
12:00 224.8 F 7.4
14:00 235.6 F 12.2
16:00 236.0 F 9.2
18:00 239.7 F 9.9
20:00 247.1 F 13.7
22:00 258.5 F 19.8
24:00 268.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.2 F 28.9
2:00 272.3 F 28.8
3:00 271.7 F 28.3
4:00 271.2 F 28.1

Min 271.2 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 202.5 F 0.0
10:00 208.0 F 0.0
15:00 205.4 F 0.0
20:00 205.3 F 0.0
25:00 206.6 F 0.0
30:00 209.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:17
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:03:59
CYCLE COMPLETE

11/07/2024 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____