

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3110

Name: RITTER 1  
Sterilizer ID: V2178198  
04/13/2023 03:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
2:00 178.5 F 2.8  
4:00 198.7 F 7.4  
6:00 218.3 F 13.1  
8:00 225.7 F 7.8  
10:00 238.4 F 12.8  
12:00 237.8 F 10.3  
14:00 236.8 F 8.2  
16:00 249.4 F 15.1  
18:00 261.3 F 21.9  
20:00 272.0 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.7 F 29.3  
2:00 271.9 F 28.6  
3:00 271.7 F 28.9  
4:00 272.7 F 29.3  
  
Min 271.4 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 223.2 F 0.0  
10:00 234.5 F 0.0  
15:00 226.5 F 0.0  
20:00 226.1 F 0.0  
25:00 226.5 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 20:35  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

04/13/2023 04:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_