
Midmark M11 - v1.0.0
Total Cycles: 4494

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2024 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 73.4 F 0.2
4:00 87.1 F 0.8
6:00 120.5 F 3.1
8:00 149.5 F 6.0
10:00 181.7 F 9.5
12:00 196.0 F 13.1
14:00 210.7 F 4.0
16:00 219.8 F 7.4
18:00 230.8 F 11.0
20:00 229.5 F 6.5
22:00 241.1 F 11.2
24:00 235.8 F 8.3
26:00 249.7 F 14.7
28:00 260.8 F 20.9
30:00 271.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 28.7
2:00 272.0 F 28.7
3:00 272.4 F 28.6
4:00 272.2 F 28.8

Min 271.6 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 228.6 F 0.0
10:00 206.7 F 0.0
15:00 204.8 F 0.0
20:00 188.9 F 0.0
25:00 241.4 F 0.0
30:00 242.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

04/26/2024 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____