
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000092
13:45 28/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:14
STERILIZING 4:37
136.4 C 226kPa 4:37
136.2 C 221kPa 5:07
136.4 C 223kPa 5:37
136.3 C 225kPa 6:07
136.3 C 225kPa 6:37
136.4 C 226kPa 7:07
136.4 C 225kPa 7:37
136.4 C 224kPa 8:07
136.4 C 224kPa 8:07
VENTING 8:07
AIR DRYING 9:04
CYCLE COMPLETE 12:29

Digital Signature #
449DDC3CA1A58BF6
