

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4261

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2024 08:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 85.7 F 0.3  
4:00 93.5 F 1.2  
6:00 114.7 F 3.3  
8:00 145.9 F 6.0  
10:00 173.2 F 9.3  
12:00 193.6 F 12.8  
14:00 209.2 F 3.6  
16:00 218.3 F 6.9  
18:00 226.1 F 10.3  
20:00 232.0 F 8.6  
22:00 237.3 F 9.8  
24:00 237.9 F 9.5  
26:00 245.6 F 12.9  
28:00 257.0 F 19.0  
30:00 267.6 F 25.4  
32:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.4 F 28.5  
2:00 272.8 F 28.9  
3:00 272.4 F 29.1  
4:00 271.9 F 28.4  
  
Min 271.6 F 28.4  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.2 F 0.0  
5:00 252.1 F 0.0  
10:00 223.7 F 0.0  
15:00 230.1 F 0.0  
20:00 231.3 F 0.0  
25:00 230.3 F 0.0  
30:00 222.5 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:04  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:10:18  
CYCLE COMPLETE

02/02/2024 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_