
Midmark M11 - v1.0.0
Total Cycles: 2447

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 181.2 F 3.3
4:00 193.5 F 8.0
6:00 213.7 F 12.3
8:00 225.9 F 8.5
10:00 238.6 F 11.0
12:00 239.7 F 11.6
14:00 238.8 F 10.2
16:00 252.6 F 17.2
18:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.7 F 29.6
2:00 271.7 F 28.8
3:00 272.6 F 29.4
4:00 271.7 F 28.8

Min 271.6 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 209.6 F 0.0
10:00 220.6 F 0.0
15:00 222.2 F 0.0
20:00 219.5 F 0.0
25:00 217.8 F 0.0
Dry Cycle Aborted

FILLING: 2:06
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 28:03
TOTAL CYCLE: 00:57:07

10/18/2022 10:37 AM
ACCEPTED: X
REJECTED:
Operator: U02