
Midmark M11 - v1.0.0
Total Cycles: 1889

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 74.4 F 0.0
4:00 85.3 F 0.0
6:00 124.8 F 0.2
8:00 152.9 F 1.2
10:00 175.5 F 3.4
12:00 195.2 F 6.2
14:00 214.7 F 9.4
16:00 225.3 F 12.9
18:00 226.4 F 6.5
20:00 237.1 F 10.7
22:00 233.7 F 7.7
24:00 243.3 F 12.4
26:00 240.3 F 10.4
28:00 252.7 F 16.7
30:00 264.1 F 23.2
32:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.9
2:00 271.4 F 28.1
3:00 272.2 F 29.0
4:00 272.3 F 28.5

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 201.5 F 0.0
10:00 208.3 F 0.0
15:00 225.1 F 0.0
20:00 248.7 F 0.0
25:00 231.3 F 0.0
30:00 228.7 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:14
CYCLE COMPLETE

05/02/2022 08:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____