

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006831  
8:49 16/01/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:05  
PRESSURIZING 4:06  
STERILIZING 4:54  
136.0 C 220kPa 4:54  
136.4 C 220kPa 5:24  
136.2 C 219kPa 5:54  
136.3 C 222kPa 6:24  
136.4 C 221kPa 6:54  
136.1 C 222kPa 7:24  
136.4 C 223kPa 7:54  
136.4 C 222kPa 8:24  
136.4 C 222kPa 8:24  
VENTING 8:24  
AIR DRYING 9:14  
CYCLE COMPLETE 13:10

-----

Digital Signature #  
CB673D2958904811  
-----