
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 003657
8:07 12/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 4:08
136.4 C 220kPa 4:08
136.4 C 223kPa 4:38
136.4 C 224kPa 5:08
136.3 C 223kPa 5:38
136.3 C 220kPa 6:08
136.2 C 221kPa 6:38
136.4 C 220kPa 7:08
136.4 C 222kPa 7:38
136.4 C 222kPa 7:38
VENTING 7:39
AIR DRYING 8:27
CYCLE COMPLETE 9:13

Digital Signature #
E1A3D8AC3DB0FBD8
