
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005488
7:49 04/05/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:07
STERILIZING 4:05
136.4 C 222kPa 4:05
136.2 C 221kPa 4:35
136.2 C 219kPa 5:05
136.3 C 222kPa 5:35
136.4 C 220kPa 6:05
136.4 C 223kPa 6:35
136.3 C 220kPa 7:05
136.1 C 221kPa 7:35
136.1 C 221kPa 7:35
VENTING 7:35
AIR DRYING 8:25
CYCLE COMPLETE 10:45

Digital Signature #
439554BDD12EA2DD
