
Midmark M11 - v1.0.5
Total Cycles: 1424

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 194.9 F 2.2
4:00 195.5 F 6.8
6:00 219.6 F 12.4
8:00 228.9 F 8.1
10:00 240.8 F 12.2
12:00 241.2 F 12.0
14:00 240.8 F 11.2
16:00 254.1 F 17.9
18:00 265.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.6
2:00 271.6 F 28.8
3:00 272.5 F 29.8
4:00 272.2 F 29.3

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 214.6 F 0.0
10:00 230.1 F 0.0
15:00 228.6 F 0.0
20:00 228.1 F 0.0
25:00 228.1 F 0.0
30:00 225.8 F 0.0

Drying Complete

FILLING: 1:45
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

08/29/2022 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____