
Midmark M11 - v1.0.0
Total Cycles: 3093

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 189.5 F 3.7
4:00 210.5 F 8.7
6:00 223.5 F 8.1
8:00 232.9 F 10.2
10:00 235.7 F 7.9
12:00 243.4 F 12.0
14:00 245.2 F 12.9
16:00 257.7 F 19.7
18:00 269.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.2 F 28.8
2:00 272.5 F 29.3
3:00 271.8 F 28.6
4:00 272.6 F 29.3

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 229.4 F 0.0
10:00 227.5 F 0.0
15:00 225.5 F 0.0
20:00 224.5 F 0.0
25:00 224.8 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

04/11/2023 03:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____