

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3160

Name: RITTER2  
Sterilizer ID: V2313228  
11/27/2023 09:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.7 F 0.0  
2:00 84.0 F 0.0  
4:00 81.8 F 0.3  
6:00 96.2 F 1.5  
8:00 137.3 F 4.1  
10:00 165.0 F 7.0  
12:00 186.5 F 10.2  
14:00 212.6 F 7.2  
16:00 217.1 F 6.5  
18:00 227.7 F 10.4  
20:00 232.3 F 8.4  
22:00 238.5 F 11.0  
24:00 234.5 F 8.7  
26:00 248.3 F 14.8  
28:00 259.0 F 20.9  
30:00 269.1 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.5  
1:00 272.3 F 29.4  
2:00 272.2 F 29.7  
3:00 272.9 F 29.9  
4:00 272.7 F 30.2

Min 271.7 F 29.4  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 238.8 F 0.0  
10:00 201.8 F 0.0  
15:00 191.1 F 0.0  
20:00 173.8 F 0.0  
25:00 170.9 F 0.0  
30:00 169.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

11/27/2023 11:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_