
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002005
13:06 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:20
STERILIZING 4:19
136.5 C 225kPa 4:19
136.2 C 224kPa 4:49
136.4 C 227kPa 5:19
136.4 C 225kPa 5:49
136.2 C 223kPa 6:19
136.4 C 227kPa 6:49
136.4 C 224kPa 7:19
136.4 C 226kPa 7:49
136.4 C 226kPa 7:49
VENTING 7:50
AIR DRYING 8:47
CYCLE COMPLETE 16:27

Digital Signature #
FB914A3E84B5C659
