

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 000637  
16:02 08/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:23  
PRESSURIZING 3:27  
STERILIZING 4:50  
136.5 C 222kPa 4:50  
136.1 C 230kPa 5:20  
136.4 C 232kPa 5:50  
136.1 C 231kPa 6:20  
136.4 C 232kPa 6:50  
136.2 C 231kPa 7:20  
136.3 C 232kPa 7:50  
136.3 C 230kPa 8:20  
136.3 C 230kPa 8:20  
VENTING 8:20  
AIR DRYING 9:13  
CYCLE COMPLETE 24:13  
-----

Digital Signature #  
B117177772C8B762  
-----