
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007414
6:48 24/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 3:02
136.4 C 222kPa 3:02
136.2 C 220kPa 3:32
136.2 C 219kPa 4:02
136.3 C 220kPa 4:32
136.4 C 223kPa 5:02
136.4 C 222kPa 5:32
136.4 C 221kPa 6:02
136.1 C 220kPa 6:32
136.1 C 220kPa 6:32
VENTING 6:32
AIR DRYING 7:22
CYCLE COMPLETE 37:22

Digital Signature #
CD1D2C62AC8FD5A7
