

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3258

Name: RITTER2  
Sterilizer ID: V2313228  
12/27/2023 05:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.0 F 0.0  
2:00 136.9 F 1.2  
4:00 158.7 F 4.2  
6:00 180.2 F 8.4  
8:00 203.4 F 13.4  
10:00 225.5 F 7.6  
12:00 236.0 F 12.7  
14:00 235.1 F 9.5  
16:00 237.8 F 9.3  
18:00 247.5 F 14.4  
20:00 260.0 F 21.3  
22:00 271.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.5 F 29.2  
2:00 271.8 F 29.5  
3:00 271.9 F 29.0  
4:00 272.2 F 29.6

Min 271.0 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.1 F 0.0  
5:00 236.2 F 0.0  
10:00 221.4 F 0.0  
15:00 186.8 F 0.0  
20:00 183.9 F 0.0  
25:00 182.5 F 0.0  
30:00 181.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:40  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

12/27/2023 06:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_