
Midmark M11 - v1.0.0
Total Cycles: 4506

Name: RITTER 1
Sterilizer ID: V2178198
04/30/2024 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.6 F 0.0
2:00 103.7 F 0.9
4:00 133.4 F 3.1
6:00 159.4 F 5.9
8:00 177.4 F 9.1
10:00 193.5 F 12.9
12:00 208.3 F 4.2
14:00 219.6 F 7.6
16:00 229.9 F 11.5
18:00 229.6 F 7.2
20:00 241.2 F 12.2
22:00 238.7 F 10.0
24:00 250.5 F 15.7
26:00 260.8 F 21.5
28:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.9 F 28.9
2:00 272.4 F 28.9
3:00 271.3 F 28.3
4:00 272.1 F 29.2

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 204.2 F 0.0
10:00 203.2 F 0.0
15:00 203.3 F 0.0
20:00 200.5 F 0.0
25:00 202.9 F 0.0
30:00 204.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:45
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:22
CYCLE COMPLETE

04/30/2024 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____