
Midmark M11 - v1.0.5
Total Cycles: 121

Name: RITTER2
Sterilizer ID: V2313228
06/21/2021 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 181.9 F 1.8
4:00 179.8 F 5.4
6:00 196.7 F 9.9
8:00 216.9 F 7.5
10:00 225.6 F 9.5
12:00 233.2 F 10.2
14:00 240.5 F 12.3
16:00 240.5 F 11.2
18:00 252.4 F 17.3
20:00 263.1 F 23.7
22:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.0
2:00 271.8 F 29.4
3:00 271.8 F 29.8
4:00 272.4 F 29.8

Min 271.2 F 29.2
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 231.9 F 0.0
10:00 230.1 F 0.0
15:00 225.8 F 0.0
20:00 223.1 F 0.0
25:00 226.3 F 0.0
Dry Cycle Aborted

FILLING: 1:45
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:24
DRYING: 28:00
TOTAL CYCLE: 00:59:30

06/21/2021 09:59 AM
ACCEPTED: X
REJECTED:
Operator: U01