
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001428
14:23 09/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:41
PRESSURIZING 6:20
STERILIZING 7:51
136.0 C 220kPa 7:51
136.5 C 223kPa 8:21
135.6 C 217kPa 8:51
136.5 C 223kPa 9:21
136.1 C 221kPa 9:51
136.4 C 222kPa 10:21
136.3 C 222kPa 10:52
136.5 C 223kPa 11:21
VENTING 11:21
AIR DRYING 12:21
CYCLE COMPLETE 21:53

Digital Signature #
9584DCB491BA4EDA
