
Midmark M11 - v1.0.5
Total Cycles: 474

Name: RITTER2
Sterilizer ID: V2313228
10/22/2021 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 193.3 F 3.7
4:00 208.2 F 9.7
6:00 224.4 F 7.0
8:00 236.5 F 11.7
10:00 238.0 F 10.1
12:00 236.5 F 9.2
14:00 251.8 F 16.5
16:00 263.8 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 271.8 F 28.8
2:00 271.8 F 29.0
3:00 272.1 F 29.3
4:00 272.3 F 29.6

Min 271.5 F 28.8
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 205.5 F 0.0
10:00 202.1 F 0.0
15:00 197.9 F 0.0
20:00 193.6 F 0.0
25:00 190.6 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 2:42
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

10/22/2021 12:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____