
Midmark M11 - v1.0.5
Total Cycles: 277

Name: RITTER2
Sterilizer ID: V2313228
08/10/2021 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.9 F 0.0
2:00 95.9 F 0.7
4:00 101.8 F 2.7
6:00 124.2 F 6.0
8:00 166.9 F 10.4
10:00 202.6 F 4.0
12:00 224.6 F 7.7
14:00 237.0 F 12.3
16:00 234.5 F 9.0
18:00 240.1 F 10.8
20:00 243.6 F 12.6
22:00 255.0 F 18.5
24:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.1 F 29.0
2:00 272.0 F 29.5
3:00 271.7 F 29.5
4:00 271.7 F 29.0

Min 271.0 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 222.6 F 0.0
10:00 233.4 F 0.0
15:00 229.2 F 0.0
Dry Cycle Aborted

FILLING: 1:02
HEATING: 25:41
STERILIZING: 4:00
VENTING: 2:56
DRYING: 19:07
TOTAL CYCLE: 00:52:46

08/10/2021 01:33 PM
ACCEPTED: X
REJECTED:
Operator: U01