
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004830
14:35 07/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:29
STERILIZING 3:30
136.5 C 220kPa 3:30
136.4 C 224kPa 4:00
136.3 C 224kPa 4:30
136.3 C 223kPa 5:00
136.3 C 223kPa 5:30
136.1 C 222kPa 6:00
136.4 C 225kPa 6:30
136.1 C 221kPa 7:00
136.1 C 221kPa 7:00
VENTING 7:00
AIR DRYING 7:49
CYCLE COMPLETE 10:36

Digital Signature #
277FD23B364EB9BF
