
Midmark M11 - v1.0.5
Total Cycles: 2075

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 182.0 F 2.2
4:00 194.7 F 6.9
6:00 214.4 F 12.5
8:00 226.1 F 8.8
10:00 235.7 F 10.1
12:00 242.9 F 13.2
14:00 243.4 F 12.5
16:00 256.8 F 19.6
18:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 273.1 F 29.8
2:00 271.9 F 29.1
3:00 272.1 F 29.8
4:00 272.9 F 29.7

Min 271.5 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 229.3 F 0.0
10:00 228.4 F 0.0
15:00 232.0 F 0.0
20:00 215.6 F 0.0
25:00 217.7 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

02/22/2023 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____