
Midmark M11 - v1.0.0
Total Cycles: 3536

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2023 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.6 F 0.0
2:00 190.9 F 2.2
4:00 201.4 F 6.6
6:00 218.5 F 11.3
8:00 224.3 F 6.1
10:00 236.0 F 10.9
12:00 233.4 F 8.1
14:00 242.8 F 11.5
16:00 246.0 F 13.2
18:00 257.8 F 19.5
20:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.4 F 29.0
2:00 271.7 F 28.2
3:00 272.3 F 29.0
4:00 272.0 F 28.5

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 212.8 F 0.0
10:00 219.4 F 0.0
15:00 214.5 F 0.0
20:00 206.3 F 0.0
25:00 203.1 F 0.0
30:00 203.8 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

07/17/2023 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____