

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2345

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2023 07:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.9 F 0.0  
2:00 84.7 F 0.0  
4:00 112.6 F 0.0  
6:00 107.3 F 0.0  
8:00 141.1 F 0.7  
10:00 173.2 F 3.8  
12:00 204.4 F 6.9  
14:00 223.6 F 10.9  
16:00 222.6 F 6.1  
18:00 229.5 F 9.4  
20:00 238.1 F 11.4  
22:00 239.7 F 11.0  
24:00 237.0 F 9.1  
26:00 249.6 F 15.3  
28:00 260.4 F 21.5  
30:00 270.6 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.4 F 29.2  
2:00 271.8 F 29.4  
3:00 272.1 F 28.9  
4:00 272.3 F 29.7

Min 271.3 F 28.7  
Max 272.4 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 232.9 F 0.0  
10:00 223.1 F 0.0  
15:00 216.5 F 0.0  
20:00 210.0 F 0.0  
25:00 228.0 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:48  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:48  
CYCLE COMPLETE

05/10/2023 08:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_