
Midmark M11 - v1.0.0
Total Cycles: 2520

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.3 F 0.0
2:00 143.9 F 0.4
4:00 179.1 F 3.8
6:00 192.1 F 7.9
8:00 206.3 F 13.2
10:00 221.9 F 6.8
12:00 232.8 F 11.7
14:00 233.4 F 8.3
16:00 241.2 F 11.1
18:00 245.0 F 12.7
20:00 256.9 F 19.2
22:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.2
2:00 271.6 F 28.6
3:00 272.7 F 29.3
4:00 271.8 F 28.5

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 226.1 F 0.0
10:00 235.7 F 0.0
15:00 236.8 F 0.0
20:00 221.9 F 0.0
25:00 216.7 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

11/07/2022 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____