
Midmark M11 - v1.0.5
Total Cycles: 283

Name: RITTER2
Sterilizer ID: V2313228
08/18/2021 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 136.3 F 1.4
4:00 151.2 F 3.9
6:00 181.2 F 7.1
8:00 196.6 F 11.1
10:00 214.0 F 5.0
12:00 221.7 F 8.5
14:00 232.5 F 12.8
16:00 234.0 F 9.7
18:00 239.1 F 10.0
20:00 245.4 F 13.8
22:00 256.3 F 19.6
24:00 266.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.8 F 30.3
2:00 271.8 F 29.6
3:00 271.3 F 29.8
4:00 272.5 F 30.2

Min 271.2 F 29.4
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 205.2 F 0.0
10:00 210.8 F 0.0
15:00 210.2 F 0.0
20:00 209.1 F 0.0
25:00 208.2 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:39
CYCLE COMPLETE

08/18/2021 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____