

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2305

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2022 08:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 74.9 F 0.1  
4:00 88.9 F 0.2  
6:00 126.6 F 0.4  
8:00 155.5 F 1.7  
10:00 185.5 F 5.6  
12:00 203.2 F 9.9  
14:00 219.6 F 8.3  
16:00 226.3 F 7.3  
18:00 236.3 F 11.3  
20:00 232.2 F 7.2  
22:00 244.6 F 13.0  
24:00 242.0 F 11.1  
26:00 255.1 F 17.8  
28:00 266.5 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.9 F 28.4  
2:00 272.4 F 28.7  
3:00 272.7 F 28.9  
4:00 272.8 F 28.8

Min 271.6 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 236.3 F 0.0  
Dry Cycle Aborted

FILLING: 0:54  
HEATING: 29:26  
STERILIZING: 4:00  
VENTING: 2:45  
DRYING: 2:20  
TOTAL CYCLE: 00:39:26

09/09/2022 09:11 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_