
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008121
10:25 02/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:42
136.3 C 220kPa 4:42
136.3 C 223kPa 5:12
136.3 C 222kPa 5:42
136.2 C 220kPa 6:12
136.4 C 224kPa 6:42
136.2 C 220kPa 7:12
136.3 C 223kPa 7:42
136.4 C 223kPa 8:12
136.4 C 223kPa 8:12
VENTING 8:12
AIR DRYING 9:05
CYCLE COMPLETE 18:39

Digital Signature #
AAB4D93C7DE224CD
