
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005472
7:44 03/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:08
136.0 C 220kPa 4:08
136.1 C 220kPa 4:38
136.3 C 222kPa 5:08
136.4 C 221kPa 5:38
136.1 C 221kPa 6:08
136.0 C 220kPa 6:38
136.3 C 222kPa 7:08
136.3 C 218kPa 7:38
136.3 C 218kPa 7:38
VENTING 7:38
AIR DRYING 8:29
CYCLE COMPLETE 8:35

Digital Signature #
124545798D0F60AD
