
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002164
9:08 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:02
STERILIZING 3:06
136.0 C 222kPa 3:06
136.4 C 226kPa 3:36
136.4 C 224kPa 4:06
136.2 C 224kPa 4:36
136.4 C 225kPa 5:06
136.4 C 225kPa 5:36
136.4 C 224kPa 6:06
136.2 C 224kPa 6:36
136.2 C 224kPa 6:36
VENTING 6:36
AIR DRYING 7:27
CYCLE COMPLETE 37:27

Digital Signature #
C4D4004A96FA2B67
